

## Instrument Security Procedures

**Model:**

Fluke 902

**Product Name:**

HVAC Clamp Meter

**Instrument Description:**

Battery-powered, true-rms indicating multimeter with a 6000-count 3 3/4-digit display.

**Memory Description:**

The Fluke 902 has one memory device: U1, micro controller, TI MSP430F427, 32kB Flash, 1024B RAM, contains operating code for the product and calibration constants.

**Memory Cleaning Instructions:**

The user controlled storage operation, other than calibration constants, is the storage of Min, Max and Hold. Min, Max and Hold is cleared when the function switch is changed.