



Aktiver FET Tastkopf | Active FET Probe

für Oszilloskope

for Oscilloscopes



TT-AF 1200

Bandwidth	DC to 1.2 GHz (-3 dB)	Ambient Operating Temp.	-10 °C to 40 °C
Attenuation Ratio	x10	Ambient Storage Temp.	-30 °C to 70 °C
Accuracy	±2 %	Ambient Operating Humidity	0 % to 85 % humidity
Input Impedance	1 MΩ, 3 pF	Ambient Storage Humidity	0 % to 85 % humidity
Rise Time	291 ps	Power Requirements	9 V battery or mains adapter
Input Voltage	±15 V (DC+peak AC)	Standard	5 VDC / 200 mA or 9 VDC / 150 mA
	+40 V (DC+peak AC)	Power Requirements	power leads
	absolute maximum voltage	Optional	(USB, Probus® or Lemo®)
Output Voltage - Swing	±1,5 V (into 50 Ω load)	Length of BNC Cable	120 cm
Output Voltage - Offset	<±5 mV	Dimensions (L x W x H)	83 mm x 19 mm x 14 mm
Adjustable output offset	-28 mV to +28 mV	Weight	200 g
Output Voltage - Noise	0,9 mVrms	Order number	12050
Output Impedance	50 Ω (for using	Delivery Content	16pcs accessory kit, mains adapter, hardcase, trimming tool, manual (de/en), calibration cert.
	50 Ω input system oscilloscope)		

Mitgeliefertes Zubehör | Accessories supplied



Adresse | Address
Fritz-Klatte-Str. 6
65933 Frankfurt am Main Germany

Kontakt | Contact
Tel.: +49 (0) 69 943335-0
Fax: +49 (0) 69 943335-55

Internet
Web: www.testec.de
E-Mail: info@testec.de