

## ***EC – DECLARATION OF CONFORMITY***

The product named hereinafter was developed, designed, and manufactured in compliance with the relevant, fundamental safety and health requirements of the listed EC directives and norms. In the event of modifications that were not authorised by us or if the product is used in a manner that is not in line with the intended purpose, this declaration will be rendered void.

*Product name:* Laboratory centrifuge

*Product type:* Sigma 1-14K

*Order number:* 10020, 10021

*Directives:*

2006/42/EC	Machinery Directive
2014/35/EU	Low Voltage Directive
2014/30/EU	EMC Directive
(EU) 2015/863	RoHS Directive

*Normes:*

- EN 61010-2-020:2017
- EN 61010-2-011:2017
- EN IEC 61000-3-2:2019
- EN 61000-3-3:2013
- EN 61326-1:2013

**Sigma Laborzentrifugen GmbH**

An der Unteren Söse 50  
37520 Osterode  
Germany

Authorised representative  
for CE matters:  
Eckhard Tödteberg

Osterode, 07/04/2020



General Manager