

## **EC – DECLARATION OF CONFORMITY**

The product named hereinafter was developed, designed, and manufactured in compliance with the relevant, fundamental safety and health requirements of the listed EC directives and norms. In the event of modifications that were not authorised by us or if the product is used in a manner that is not in line with the intended purpose, this declaration will be rendered void.

*Product name:* Laboratory centrifuge

*Product type:* Sigma 4-16KS, Sigma 4-16KHS

*Order number:* 10385, 10386, 10387, 10388, 10389,  
90971, 90978

*Directives:* 2006/42/EC Machinery Directive  
2014/35/EU Low Voltage Directive  
2014/30/EU EMC Directive  
(EU) 2015/863 RoHS Directive

*Normes:* EN 61010-2-020:2017  
EN 61010-2-011:2017  
EN IEC 61000-3-2:2019  
EN 61000-3-3:2013  
EN 61326-1:2013

### **Sigma Laborzentrifugen GmbH**

An der Unteren Söse 50  
37520 Osterode  
Germany

Authorised representative  
for CE matters:  
Eckhard Tödteberg

Osterode, 07/04/2020



General Manager