


Beta spectroscopy (Item No.: P2523200)

Curricular Relevance



Difficulty



Difficult

Preparation Time



1 Hour

Execution Time



2 Hours

Recommended Group Size



2 Students

Additional Requirements:

Experiment Variations:

Keywords:

β^- decay, β^+ decay, electron capture, neutrino, positron, decay diagram, decay energy, resting energy, relativistic Lorentz equation

Overview

Short description

Principle

The radiation of β^- -unstable atomic nuclei is selected on the basis of its pulses in a magnetic transverse field, using a diaphragm system. The relationship between coil current and particle energy is determined for calibration of the spectrometer and the decay energy of the β^- -transition is obtained in each case from the β^- -spectra.



Fig. 1: Experimental set-up for determining inductance from the resonant frequency of an oscillatory circuit.

Equipment

Position No.	Material	Order No.	Quantity
1	Beta-spectroscope	09104-00	1
2	Iron core, solid, 25 mm long	06490-01	1
3	Iron core, short, laminated	06500-00	1
4	Iron core, U-shaped, laminated	06501-00	1
5	Clamping device	06506-00	1
6	Coil, 600 turns	06514-01	1
7	Radioactive source Na-22, 74 kBq	09047-52	1
8	Radioactive source Sr-90, 74 kBq	09047-53	1
9	Geiger-Mueller counter tube, 15 mm (type B)	09005-00	1
10	Geiger-Müller-Counter	13606-99	1
11	PHYWE power supply, universal DC: 0...18 V, 0...5 A / AC: 2/4/6/8/10/12/15 V, 5 A	13500-93	1
12	Digital multimeter 2005	07129-00	1
13	Teslameter, digital	13610-93	1
14	Hall probe, tangential, protection cap	13610-02	1
15	Connecting cord, 32 A, 500 mm, red	07361-01	1
16	Connecting cord, 32 A, 500 mm, blue	07361-04	2

Tasks

1. Energy calibration of the magnetic spectrometer.
2. Measurement of the β -spectra of ^{90}Sr and ^{22}Na .
3. Determination of the decay energy of the two isotopes.

Set-up and procedure

The experiment is set up as shown in Fig. 1, initially without the source. The iron components must be firmly fixed in position so as to ensure a satisfactory and constant flow of forces. The zero point has to be adjusted on the magnetic field measuring instrument, before the tangential Hall probe is introduced through the lateral opening. The relationship between coil current and magnetic flux density is determined. Measurements are carried out in both directions of the magnetic field (see Fig. 2).




Fig. 2: Calibration of the spectrometer: relationship between coil current and selected particle energy.

The source and the counter tube are inserted, the Geiger- Müller counter connected and, after establishing the correct direction of the magnetic field, the counting rate per 10 s measurement period is determined at different field strengths. The measurement is recorded for both isotopes, determining in each case the zero effect with the source but in opposite directions of the field.

Theory and evaluation

β -particles are selected in the β -spectroscopie on the basis of their energy by obliging them to follow a fixed orbit using diaphragms in a homogeneous magnetic field (Fig. 3 and 4).




Fig. 3: Diaphragm system and magnetically deflected β -particle; r = orbital radius.




Fig. 4: β -spectroscopie: 1) non magnetisable wall; 2) specimen entrance; 3) iris; 4) entrance for tangential Hall probe; 5) counter tube holder; 6) iris.

In this orbit the Lorentz force, due to the magnetic cross-field, and the centrifugal force are in equilibrium:

$$e \cdot v \cdot m = \frac{m \cdot v^2}{r} .$$

This yields the following expression for the momentum:

$$p = m \cdot v = e \cdot B \cdot r$$

The equation for relativistic particles with the momentum is then

$$\frac{E^2}{c^2} = p^2 + m_0^2 c^2 ,$$

in which E denotes the total energy of the particles, made up of the kinetic energy E_{kin} and the resting energy $m_0 c^2$


$$E = E_{\text{kin}} + m^0 \cdot c^2 .$$

The kinetic energy is accordingly.


$$E_{\text{kin}} = \sqrt{(eBrc)^2 + m_0^2 c^4} - m_0 c^2 .$$

With a given orbital radius of $r = 50 \text{ mm}$, it is possible to fix a specific particle energy for each magnetic field strength and for each current strength.

The process of β^- -decay in the atomic nucleus results in the conversion of a neutron n into a proton p and an electron e^- , which leaves the nucleus, and into an antineutrino $\bar{\nu}$ which is difficult to detect. The complete decay equation then runs as follows:



β^+ -decay causes the occurrence of a positron, in which case the decay equation will be:



The decay energy E_Z is released during the conversion (Figs. 5 and 6).




Fig. 5: Decay diagram of ^{90}Sr




Fig. 6: Decay diagram of ^{22}Na .

Since the neutrino carries with it a proportion of the decay energy, the magnitude of which cannot be determined, a continuous energy distribution occurs in which all the energy values from 0 to E_Z occur.

A further characteristic of the β -spectrum is its most frequent energy E_h , which will always be one third of the maximum energy: E_Z

$$E_h = 1/3 E_Z .$$

The most frequent energy E_h can be determined with a very much greater accuracy than the maximum energy E_Z , since at the upper end the spectrum passes with a flat slope into the zero effect.

Fig. 7 yields on the basis of the ^{90}Sr energy spectrum a maximum energy of about $E_Z = 2500$ keV and the following value for the most frequent energy:

$$E_h = 740 \pm 30 \text{ keV} .$$





Fig. 8 yields on the basis of the ^{22}Na energy spectrum a maximum energy of about $E_Z = 900$ keV and the following value for the most frequent energy:

$$E_h = 210 \pm 30 \text{ keV} .$$



Notes

1. The direction of the magnetic field will vary in some of the experiments, since the radiation particles have a different sign.
2. More careful examination reveals that the scaling of the spectral intensities is not constant in the measured spectra for the following reasons:
 - (a) the interval width per energy measurement is a momentum window and not an energy window.
 - (b) the probability of response of the counter tube and the resolution of the spectrometer are energy-dependent. The spectra are as a result distorted, although this does not affect their predictive value in relation to the energies.
3. The greatly intensified background in the ^{22}Na spectrum is attributable to the 511 keV annihilation radiation.

4. The spectra are made up of several fractions in both cases. In conjunction with the limited resolution of the spectrometer, this results in a systematic measurement error. The most frequent particle energy is in agreement with the theoretical values for both isotopes within the limits of measurement accuracy.